Dispensing guide

Measurements required						
	Rx	Pupil distance	Fitting height	Corridor length	Lens markings	
SuperDigital	\checkmark	\checkmark	\checkmark	\checkmark	D1, D2, D3	

Fitting guide	
Minimum fitting height	14 mm

This allows freedom to choose wider range of frame styles and lens shapes.

Clear		1.5, 1.6, 1.67, 1.74
Reactions		1.5, 1.6, 1.67
Polarising:	Grey	1.5, 1.6, 1.67
	Brown	1.5, 1.6, 1.67
	Green	1.5

Technical information

Optimised reading insets:

The position of the reading inset varies and is Optimised for the wearer's specific prescription. Average powers for Pantoscopic Tilt 9 degrees, wrap angle 6.5 degrees and BVD 12 mm are used.

Fixed corridor length:

The definition of corridor length (CL) for these specific designs is the distance from the fitting cross to the centre of the reading area. 3 fixed CLs to cover a range of fitting heights have been pre-selected:

Lens A

Minimum fitting height 14-17mm 14mm corridor length

Lens B Minimum fitting height 18-20mm 15mm corridor length

Fitting position:

The position of the fitting cross is 4mm above the prism reference point.



Lens C Minimum fitting height 21-23mm 18mm corridor length